

Acetyl CoA Assay Kit ab87546

★★★★★ [1 Abreviews](#) [57 References](#) [画像数 1](#)

製品の概要

製品名	Acetyl CoA Assay Kit
検出方法	Fluorescent
サンプルの種類	Cell Lysate, Tissue Lysate
アッセイタイプ	Quantitative
検出感度	> 0.01 nmol/well
検出範囲	0.01 nmol/well - 0.5 nmol/well
全工程の試験時間	0h 20m
製品の概要	<p>Acetyl CoA Assay Kit ab87546 is a highly sensitive assay for quantifying Acetyl CoA level in biological samples.</p> <p>In the Acetyl CoA assay protocol, free CoA is quenched then Acetyl CoA is converted to CoA. The CoA is reacted to form NADH which interacts with a probe to generate fluorescence (Ex=535/Em=587 nm).</p> <p>The assay can detect 10 to 500 pmol of Acetyl CoA (with detection limit ~0.4 μM).</p> <p>Acetyl CoA assay protocol summary:</p> <ul style="list-style-type: none">- add samples and standards to wells- add CoA quencher to wells to remove background from free CoA and succ-CoA and incubate at room temp for 5 min- add quencher remover and incubate at room temp for 5 min- add reaction mix and incubate for 10 min at 37°C- analyze with microplate reader
特記事項	<p>This product is manufactured by BioVision, an Abcam company and was previously called K317 PicoProbe™ Acetyl-CoA Fluorometric Assay Kit. K317-100 is the same size as the 100 test size of ab87546.</p> <p>Acetyl CoA is a central molecule of metabolism. It carries acetate, used in the build-up and breakdown of larger molecules.</p>
試験プラットフォーム	Microplate reader


製品の特性

保存方法 Store at -20°C. Please refer to protocols.

内容	100 tests	2000 tests
Acetyl CoA Enzyme Mix	1 x 500µl	20 x 500µl
Acetyl CoA Standard	1 vial	20 vials
Acetyl CoA Substrate Mix	1 vial	20 vials
Assay Buffer XXII	1 x 25ml	20 x 25ml
CoA Quencher	1 x 1ml	20 x 1ml
Converter Enzyme III	1 x 100µl	20 x 100µl
PicoProbe II	1 x 200µl	20 x 200µl
Quench Remover I	1 vial	20 vials

関連性 Acetyl CoA (AcCoA) is the metabolic intermediate which transfers carbon atoms to the TCA cycle for energy production and is the link between fat and carbohydrate metabolism. When fatty acid levels are high, AcCoA levels rise beyond the energy requirements of the organism and the excess is converted to the ketone bodies acetone and β -hydroxybutyrate. AcCoA also contributes acetate to histone acetylation and isoprenoid synthesis.

画像



Induction of acetyl-CoA in mitochondria by palmitate. Mitochondria were isolated from cells after palmitate treatment (300 µM) at 1, 4 and 16 h, and then used in the concentration assay of acetyl-CoA.

The acetyl-CoA content was determined immediately after isolation of the mitochondria using ab87546. Briefly, acetyl-CoA standard curve was made in the range of 0–100 pM and the correlation coefficient was 0.990 or higher. Protein was removed in the sample using the perchloric acid protocol and the supernatant was neutralized with 3 M KHCO₃. The CoASH Quencher and Quencher remover were added into the sample to correct the background generated by free CoASH and succ-CoA. The sample was then diluted with the reaction mix, and the fluorescence signal was measured at Ex/Em = 535/589 nm with Spectra max Gemini XPS (Molecular Devices, Sunnyvale, CA). The relative acetyl-CoA concentration was normalized with the mitochondrial protein.

Please note: All products are "FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES"

Our Abpromise to you: Quality guaranteed and expert technical support

- Replacement or refund for products not performing as stated on the datasheet
- Valid for 12 months from date of delivery
- Response to your inquiry within 24 hours

- We provide support in Chinese, English, French, German, Japanese and Spanish
- Extensive multi-media technical resources to help you
- We investigate all quality concerns to ensure our products perform to the highest standards

If the product does not perform as described on this datasheet, we will offer a refund or replacement. For full details of the Abpromise, please visit <https://www.abcam.co.jp/abpromise> or contact our technical team.

Terms and conditions

- Guarantee only valid for products bought direct from Abcam or one of our authorized distributors